



# Certificate of Compliance

**Certificate:** 80063651

**Master Contract:** 301694

**Project:** 80063651

**Date Issued:** 2020-12-18

**Issued To:** Eternity Medical Equipment Manufactural Inc  
103-19099 25th Ave  
Surrey, British Columbia, V3Z 3V2  
Canada

**Attention:** Jeffrey Wang

*The products listed below are eligible to bear the CSA Mark shown*

**Issued by:** Wendy Zhu  
Wendy Zhu, Ph. D., P. Eng.



## **PRODUCTS**

CLASS - C720402 - OCCUPATIONAL HEALTH & SAFETY Medical Non-Powered Air-Purifying Respirators

Medical Respirator Model ECAN95, single-use and disposable, 95PFE-L2

## **APPLICABLE REQUIREMENTS**

- The Code of Federal Regulation 42 CFR Part 84, Approval of Respiratory Protective Devices, Subpart K (Nonpowered Particulate Respirators).
- The Code of Federal Regulations, Title 16- Commercial Practices, Part 1610, Standard for the Flammability of Clothing Textiles
- ASTM F1862-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood
- CSA TechNote TN202, Filtering Facepiece Respirator Head Strap / Harness Assembly Strength Requirements and Test Procedure



## *Supplement to Certificate of Compliance*

**Certificate:** 80063651

**Master Contract:** 301694

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80063651	2020-12-18	Original certification of Medical Respirator Model ECAN95, single-use and disposable, 95PFE-L2. The Code of Federal Regulation 42 CFR Part 84, Subpart K The Code of Federal Regulations, Title 16- Commercial Practices, ASTM F1862-17 CSA TechNote TN202